

# Cek Plagiasi Windra Telkomnika

1

*by* Windra .

---

**Submission date:** 29-Jan-2019 10:03AM (UTC+0700)

**Submission ID:** 1069886615

**File name:** Telkomnika1.pdf (161.09K)

**Word count:** 3356

**Character count:** 17826

## Compressed Sensing for Thoracic MRI with Partial Random Circulant Matrices

Windra Swastika<sup>1</sup>, Hideaki Haneishi<sup>2</sup>

<sup>1</sup>Informatics Engineering Study Program, Ma Chung University, Malang, Indonesia  
Villa Puncak Tidar N/01, Malang, Phone: +62-341-550171

<sup>2</sup>Research Center for Frontier Medical Engineering, Chiba University, Chiba, Japan  
1-33, Yayoi-cho, Inage-ku, Chiba 263-8522  
e-mail: windra.swastika@machung.ac.id

### Abstrak

Penggunaan matriks sirkulan sebagai matriks pengindera pada skema penginderaan kompresif (CS) telah banyak digunakan untuk mengatasi keterbatasan matriks random atau Fourier parsial. Disamping mengurangi kompleksitas komputasi, penggunaan matriks sirkulan untuk citra magnetic resonance (MR) dapat mempermudah implementasi perangkat keras. Pada makalah ini dijabarkan tentang simulasi penginderaan kompresif untuk citra MR pada organ pernafasan menggunakan matriks sirkulan sebagai matriks pengindera. Rekonstruksi dari tiga jenis citra MR dibandingkan berdasarkan jumlah sample, jumlah iterasi, dan kualitas rekonstruksi menggunakan signal to noise ratio (SNR). Dari hasil simulasi menunjukkan bahwa matriks sirkulan dapat difungsikan dengan eifisien untuk merekonstruksi citra MR, khususnya untuk citra MR yang tipis pada kawasan spasial.

**Kata kunci:** penginderaan kompresif, citra MR, organ pernafasan, matriks circulant

### Abstract

The use of circulant matrix as the sensing matrix in compressed sensing (CS) scheme has recently been proposed to overcome the limitation of random or partial Fourier matrices. Aside from reducing computational complexity, the use of circulant matrix for magnetic resonance (MR) image offers the feasibility in hardware implementations. This paper presents the simulation of compressed sensing for thoracic MR imaging with circulant matrix as the sensing matrix. The comparisons of reconstruction of three different type MR images using circulant matrix are investigated in term of number of samples, number of iteration and signal to noise ratio (SNR). The simulation results showed that circulant matrix works efficiently for encoding the MR image of respiratory organ, especially for smooth and sparse image in spatial domain.

**Keywords:** compressive sensing, MR image, respiratory organ, circulant matrix

7  
Copyright © 2012 Universitas Ahmad Dahlan. All rights reserved.

### 1. Introduction

Magnetic resonance imaging (MRI) is one of the alternatives for intra-cardiac visualization. Compared with other imaging modalities such computed tomography (CT), MRI is relatively safe since it does not using ionizing radiation (x-rays). It is also able to obtain sequential 2D slices or 3D volumes with high spatio-temporal resolution. Despite all the advantages of MRI, some major problems are inevitable. Some issues and problems of brain MRI such as segmentation, localization, correction, and classification are discussed in [1]. Long scan time in data acquisition is also found to be a major problem in MRI, especially to capture thoracic or respiratory organ. Thoracic MR image provides representation of heart, valves and major vessels anatomy and function. It is also useful to diagnose any cardiovascular problems. Ideally, physicians want to be able to observe real-time motion of respiratory organ. Unfortunately, it is impossible to get real-time motion using current MR scanner. Moreover, acquiring MR image of respiratory organ can lead to quality image degradation due to the motion during inhale and exhale of the patient. Breath holding during acquisition is one of the methods to reduce image artifact. An overview of reducing motion artifacts principles such as retrospective gating or projection reconstruction were discussed in [4]. Hardware based

approach to reduce imaging time using parallel data acquisition (P-MRI) method is proposed in [6]. Clearly, the needs to reduce imaging time which yields high quality image is the holy grail in MR imaging technology.

Compressed Sensing (CS) is a relatively new method introduced by Donoho and Candes [2-3] and expected to answer one of the issue in MR image acquisition, fast imaging time acquisition without losing its quality. It is built upon the fact that a signal (image) has a sparse representation in a priori known basis (or compressible). A signal  $u$  with length  $n$  can be transformed using an orthogonal basis  $\Psi$  (e.g. Fourier or Wavelet basis) such that  $\Psi \cdot \bar{u} = \bar{x}$ . Here  $\bar{x}$  can be seen as a sparse representation of signal  $\bar{u}$ , and it has  $k$  numbers of nonzeros ( $k$ -sparse), where  $k < n$ . Once known that vector  $\bar{x}$  is sparse or compressible, it can be compressed by projecting it onto a measurement matrix ( $\Phi$ ). Matrix  $\Phi$  is an  $m \times n$  matrix where the number of its row is much smaller than the size of the signal ( $m \ll n$ ). The projection is simply the inner product of  $\Phi$  and vector  $\bar{u}$  ( $\bar{y} = \Phi \bar{u}$ ). Since  $m \ll n$ , the equation  $\bar{y} = \Phi \bar{u}$  leads to an ill-posed problem (underdetermined system) and it is impossible to recover  $\bar{u}$  (sparse representation of original signal  $\bar{x}$ ) from  $\bar{y}$ . However, the sparsity of  $\bar{u}$  changes the impossible recovery into a perfect recovery of  $\bar{u}$ . Another condition needs to be satisfied to reconstruct  $\bar{u}$  from  $\bar{y}$  is restricted isometry property (RIP) of matrix  $\Phi$  which is obeyed by many types of matrices (e.g. Random Gaussian, Bernoulli, Partial Fourier Matrices) [7]. The reconstruction process can be done by greedy method (fast, but requires more measurements) or solving convex optimization problem (slow but requires less measurement data).

Lustig et al in [5] showed that MR image is suitable for CS implementation since it has sparse representation in certain transform domain. The DCT and Wavelet transform have good performance of recovering brain and angiogram MR images with 5-10% coefficients. Moreover, wavelet transform also proven to be effective for real-time data transmission for real image (JPEG2000) [13] showing that both medical and real image are actually sparse in when transformed into certain domain. But the success of reconstruction does not only merely depend on the sparse representation. Another factor is to satisfy RIP, which is a tool for analyzing the performance how efficient a matrix measures sparse signal [8]. Recent widely measurement matrices that satisfy RIP with high probability is independent and identically distributed (i.i.d) random Gaussian matrix. Although Gaussian matrix provides incoherence with any sparse signal and the number of measurements for signal recovery is minimal, it requires huge memory storage and high complexity computation. Using Gaussian matrix as sensing matrix to acquire a 256x256 image with 50% sampling will require nearly gigabyte space and giga-flop operation, leads to impractical application of CS scheme for MR device. Structurally random matrices, such as Toeplitz or Circulant matrix have been shown to be fast computable and satisfy RIP property as well for almost all orthonormal matrices [9]. But, the efficiency using Circulant matrix to reconstruct MR images have not been known.

In this paper, we investigated the efficiency of using circulant matrix as sensing matrix for MR Image of respiratory organ compared to partial Fourier matrix. The reconstruction is done via conventional  $\ell_1$ -norm minimization.

The contents of this paper are organized as follows: in Section II we review the basic theory of compressed sensing and MR image of respiratory organ motion. In Section III, we discuss the methodology we use in the experiment and also the simulation results. We conclude this paper in Section IV.

## 2. Basic Theory

There are three stages in CS scheme: encoding, sensing and decoding (reconstruction). Encoding is a stage of sparsifying input signal (or image) using certain basis representation (such as Fourier or wavelet). Sensing phase is a process to measure the sparse signal representation and reducing its dimension using sensing matrix and at last, decoding is a process to reconstruct the sensing signal.

This section is divided into two parts. First part discusses about basic CS theory and its reconstruction using Total Variation minimization. Second part reviews partial random circulant matrix as sensing matrix.

## 2.1. CS with Total Variation Minimization

The encoding phase starts with an input of 1-D discrete signal  $\bar{u}$  with length  $n$  that can be represented in term of an orthonormal basis  $\{\psi_i\}_{i=1}^n$  as:

$$\bar{x} = \sum_{i=1}^n \psi_i \bar{u}_i = \Psi \bar{u} \quad (1)$$

Vector  $\bar{x}$  is a sparse representation of vector  $\bar{u}$  and has  $k$  non-zeros coefficients. We say that vector  $\bar{x}$  is highly compressible if  $k \ll n$ , or in other word, non-zeros or significant coefficients much smaller than length of vector  $\bar{u}$ .

The sensing phase of CS needs measurement matrix for projecting vector  $\bar{x}$  into smaller dimension. Let  $\Phi$  is a measurements matrix with  $m \times n$  dimension (where  $k < m \ll n$ ), the linear projection of sparse vector  $\bar{x}$  is given by:

$$\bar{y} = \Phi \bar{x} = \Phi \Psi \bar{u} = A \bar{u} \quad (2)$$

The measurement matrix  $\Phi$  has to be independent of  $\bar{x}$  (non-adaptive/fixed). From (2) we have vector of measurements  $\bar{y}$  with length  $m$ .

The decoding stage is a stage to reconstruct  $\bar{x}$  from  $\bar{y}$  and matrix  $\Phi$ . Since there are  $m$  equation and  $n$  unknowns to be solved, the reconstruction becomes an ill-posed problem (underdetermined system) with infinitely many solutions. CS theory [1, 2] states that vector  $\bar{x}$  can be recovered by solving convex problem. As long as  $\bar{y}$  is highly sparse, finding the sparsest solution will be the "best guest" to recover vector  $\bar{x}$ .

$$\min \|x\|_0 \text{ subject to } Ax = y \quad (3)$$

Unfortunately, (3) is a NP-hard problem and computationally impractical. An effective way to recover  $\bar{x}$  is to solve  $\ell_1$ -norm instead of  $\ell_0$ -norm. This method is also known as Basis-pursuit.

$$\min \|x\|_1 \text{ subject to } Ax = y \quad (4)$$

By solving  $\ell_1$ -norm optimization problem, vector  $\bar{x}$  can be faithfully recovered. Another assumption in decoding phase is the number of measurements,  $m \geq ck \log n$ , for some small  $c$  where the measurements matrix is Gaussian or  $m \geq k \log n$  where the measurements matrix is random partial Fourier.

In practical application, MR image acquisition collects data in the frequency domain ( $k$ -space). Lustig *et al* in [4] proposed a reconstruction model for MR Image. Let  $\mathcal{F}$  be the undersampled Fourier transform, the reconstruction is given by:

$$\min \|\Psi \bar{u}\|_1 \text{ subject to } \|\mathcal{F} \bar{u} - y\|_2 < \epsilon \quad (5)$$

where  $\epsilon$  represents as threshold parameter for noise level. Total Variation (TV) as the sum of the absolute variations in the image can be also employed. The reconstruction model in equation (5) can be written as:

$$\min \|\Psi \bar{u}\|_1 + \alpha TV(\bar{u}) \text{ subject to } \|\mathcal{F} \bar{u} - y\|_2 < \epsilon \quad (6)$$

Here,  $\alpha$  is a positive parameter. Better reconstruction in MR image from a small number of Fourier coefficients is achieved by additional TV.

## 2.2. Partial Random Circulant Matrix as Sensing Matrix

In equation (2),  $A$  is called sensing matrix. Sensing matrix plays an important role in decoding (reconstruction) stage. Recent results show that stable reconstruction from  $k$ -sparse and compressible signal can be achieved when the sensing matrix satisfies *restricted isometry property* (RIP) [7]. Current widely used sensing matrices are random matrices and partial

Fourier. Random matrices (such as Gaussian or Bernoulli random matrices) has optimal performance but requires big memory storage and complexity computation while partial Fourier enables a fast computation of Fast Fourier Transform (FFT) but lack of universality (only incoherence with sparse signal in time domain but not for smooth signal).

Circulant matrices are orthonormal matrices that are nearly incoherence with other orthonormal matrices [10], which provide universality with many types of signal. It is formed by starting with a vector with  $N$  components. The subsequent rows are acquired by shifting the previous row to the right.

Given a vector  $c = (c_0, c_1, c_2, \dots, c_{N-1})^T \in \mathbb{R}^N$ , and  $\phi = \phi(c) \in \mathbb{R}^{N \times N}$ , the general form of circulant matrices are given by:

$$\phi(c) = \begin{bmatrix} c_0 & c_1 & \dots & c_{N-1} \\ c_{N-1} & c_0 & \dots & c_{N-2} \\ \vdots & \vdots & \ddots & \vdots \\ c_1 & c_2 & \dots & c_0 \end{bmatrix} \quad (7)$$

Consider an index set  $\theta \subset \{1, \dots, n\}$  as a subset of randomly locations between  $(1 \dots n)$  with cardinality  $m$ . The rows of  $\phi(c)$  can be restricted by the element of  $\theta$  and formed  $\phi(c)_\theta$  as partial circulant matrix which contains  $m$  number of rows.

A new paradigm of *random filter* was proposed by Tropp in [11] to captures signal  $s$  by convolution operation of random-tap FIR filter  $h$ , followed by subsampling. Figure 1 shows a block diagram of the measurement process through random filtering.




Figure 1. Block diagram of measurement process through random filtering

Using this paradigm, equation (2) can be written as:

$$\bar{y} = D_1(c * \bar{u}) \quad (8)$$

where  $D_1$  is subsample of convolution operation  $(c * \bar{u})$ . Using the FFT to implement the convolution, equation (8) can be expressed as:

$$\bar{y} = D_1(\mathcal{F}^{-1}\Sigma\mathcal{F}) * \bar{u} \quad (9)$$

where  $\mathcal{F}$  and  $\mathcal{F}^{-1}$  is Fourier matrix and its adjoint.  $\Sigma$  is diagonal matrix of  $c \in \mathbb{R}^N$  Fourier transformation. The theoretical analysis of restricted isometries for partial random circulant matrices is given by Rauhut *et al* in [10].

### 3. Results and Analysis

To demonstrate the effectiveness of partial random circulant matrix in MR image of respiratory organ, we used three models of MR images of respiratory organ from different patients. The MR images were acquired using a 1.5T Achieva Nova-Dual (Philips Medical Systems, Best, NL) whole-body scanner with a 16ch SENSE TORSO XL Coil. The size of each image is 256x256.

Experiments were performed under Windows 7 Professional running on Dell PC Desktop with an Intel® Core™ 2 Quad CPU, 2.66GHz and 4GB RAM. The reconstruction model we used in the experiments is fast algorithm proposed in [12]. This algorithm is implemented in a MatLab program, RecPC [15]. Three original images are shown in Figure 2a, 2b, and 2c respectively. Each image represents different variation of respiratory organ. Figure 2a represents front side of respiratory organ, Figure 2b represents left side and Figure 2c represents back side.




Figure 2. Original MR images of respiratory organ. (a) Front side; (b) Left side; (c) Back side

The visual result of the reconstruction using partial random circulant matrix for image in Figure 2a, 2b and 2c are shown in Figure 3, 4 and 5 respectively. As shown in Figure 3, at the sample ratio of 10% and 25%, the reconstruction images still suffer from aliasing artifact due to the sharpness original image.




Figure 3. Result of image reconstruction, original image in Figure (2a)  
(a) 10% samples (12.4dB); (b) 25% samples (16.7dB); (c) 50% samples (22.1dB)

For reconstruction of Figure 2b and 2c (as shown in Figure 4, 5), the aliasing artifact are less at sample ratio of 25% and has better quality at sample ratio of 50% compared to Figure 3. The reconstruction works better for sparse and smooth image as in Figure 2b and 2c.

Table 1 shows the comparison of each image based on number of samples, number of iteration and reconstruction quality (SNR). The best reconstruction is gained by 50% sampling of left side MR image of respiratory organ, as the sparsest in spatial among the images.

Table 1. Comparison of three different image reconstructions in term of number of samples, number of iteration and image quality

# of sample	Fig 2a (Front side)		Fig 2b (Left side)		Fig 2c (Back side)	
	# itr	SNR (dB)	# itr.	SNR (dB)	# itr	SNR (dB)
50%	140	22.1	100	29.1	90	28.4
25%	280	16.7	190	20.3	160	22.1
10%	560	12.4	430	14.4	360	17.0




Figure 4. Result of image reconstruction, original image in Figure (2b).  
(a) 10% sampling (14.4dB); (b) 25% sampling (20.3dB); (c) 50% sampling (29.1dB)




Figure 5. Result of image reconstruction, original image in Figure (2c).  
(a) 10% sampling (17.0dB); (b) 25% sampling (22.1dB); (c) 50% sampling (28.4dB)

As shown in [3], exact reconstruction of MR Images using CS scheme is achievable by employing random partial Fourier matrix. To validate our results, we compared the quality of MR Images reconstruction using circulant matrix and partial Fourier matrix. The number of sampling we picked for the comparison is 35%, 50% and 75%. This is due to the unstable reconstruction using random partial Fourier matrix when the number of sampling is below 35%. Table 2 shows the comparison of image reconstruction using circulant matrix (CM) and partial Fourier matrix (PFM) in term of SNR (dB).

As shown in Table 2, MR images reconstruction using random partial circulant matrix as sensing matrix has better SNR compared with reconstruction using partial Fourier matrix. Table 3 showed that the time required to reconstruct the image using Circulant Matrix with lower sampling has better speed compared with random partial Fourier matrix (35% and 50% sampling). In the other side, when the number of sampling is higher (75%), shorter

reconstruction time was achieved by employing partial Fourier matrix, yet the reconstruction quality (SNR) is far below the circulant matrix.

Table 2. SNR (dB) comparison using Circulant Matrix (CM) and partial Fourier matrix (PFM)

# of sampling	Fig 2a (Front side)		Fig 2b (Left side)		Fig 2c (Back side)	
	CM	PFM	CM	PFM	CM	PFM
75%	28.2	15.9	42.5	31.5	36.8	13.4
50%	21.8	10.8	29.1	23.2	28.5	8.6
35%	18.1	6.1	23.3	16.7	24.7	1.7

Table 3. Reconstruction time (sec) comparison using Circulant Matrix (CM) and partial Fourier matrix (PFM)

# of sampling	Fig 2a (Front side)		Fig 2b (Left side)		Fig 2c (Back side)	
	CM	PFM	CM	PFM	CM	PFM
75%	1.6	1.4	1.5	1.2	1.3	1.1
50%	1.6	1.7	1.4	1.5	1.6	1.4
35%	1.4	1.9	1.6	2	1.4	1.6

#### 4. Conclusion

We applied CS using partial random circulant matrix as sensing matrix for MR image of respiratory organ. The comparison of three different MR image of respiratory organ using partial random circulant matrix as sensing matrix has been performed. It has been found that partial random circulant matrix can efficiently reconstruct MR image of respiratory organ especially for sparse image in spatial domain. As partial random circulant matrix has better image quality and faster reconstruction time when using lower sampling compared with partial Fourier matrix, its implementation for MR scanner is very promising. Further study can be carried out by combining the CS scheme using partial random circulant matrix with particular medical image techniques such as segmentation [14].

#### References

- [1] N Yudistira, D Daman. Issues and Problems in Brain Magnetic Resonance Imaging: An Overview. *TELKOMNIKA*. 2008; 6(1): 57-64.
- [2] DL Donoho. Compressed Sensing. *IEEE Transactions on Information Theory*. 2006; 52 (4): 1289–1306
- [3] E Candés, J Romberg, T Tao. Robust uncertainty principles: exact signal reconstruction from highly incomplete frequency information. *Information Theory IEEE Transactions*. 2006; 52(2): 489-509.
- [4] R Van de Walle, I Lemahieu, E Achten. Magnetic Resonance Imaging and the Reduction of Motion Artifacts: Review of the Principles. *Technology and Health Care*. 1997; 5(6):419-435.
- [5] M Lustig, D Donoho, JM Pauly. Sparse MRI: The application of compressed sensing for rapid MR imaging. *Magn. Reson. Med*. 2007; 58(6): 1182-1195.
- [6] SKP Pruessmann. Encoding and reconstruction in parallel MRI. *NMR in Biomedicine*. 2006; 19(3): 288–299.
- [7] E Candés, T Tao. Decoding by linear programming. *IEEE Transaction on Information Theory*. 2005; 51: 4203-4215.
- [8] E Candés, JK Romberg, T Tao. Stable signal recovery from incomplete and inaccurate measurements. *Comm. Pure Appl. Math*. 2006; 59(8):1207–1223.
- [9] TT Do, TD Tran, L Gan. *Fast compressive sampling with structurally random matrices*. Proc. IEEE Int. Conf. on Acoustics, Speech, and Signal Processing. 2008: 3369-3372.
- [10] H Rauhut, J Romberg, J Tropp. Restricted isometries for partial random circulant matrices. Arxiv preprint. 2010; arXiv:1010.1847
- [11] J Tropp, M Wakin, M Duarte, D Baron, R Baraniuk. Random filters for compressive sampling and reconstruction. Int. Conf. Acoustics, Speech, and Signal Processing. 2006; 3: 872–875.
- [12] W Yin, SP Morgan, J Yang, Y Zhang. Practical compressive sensing with toeplitz and circulant matrices. 2010; *Rice University CAAM Technical Report TR10-01*.
- [13] M Nasri, A Helali, H Sghaier, H Maaref. Efficient JPEG 2000 Image Compression Scheme for Multihop Wireless Networks. *TELKOMNIKA*. 2011; 9(2): 311-318.



- 
- [14] A Alfiansyah. A Unified Energy Approach for B-Spline Snake in Medical Image Segmentation. *TELKOMNIKA*. 2010; 8(2): 175-186.
- [15] <http://www.caam.rice.edu/~optimization/L1/RecPC/>. Practical Compressive Sensing with Toeplitz and Circulant Matrices. Last access: Nov 24, 2011; 14:27pm.

# Cek Plagiasi Windra Telkomnika 1

## ORIGINALITY REPORT

18%

SIMILARITY INDEX

12%

INTERNET SOURCES

9%

PUBLICATIONS

3%

STUDENT PAPERS

## PRIMARY SOURCES

1	<a href="http://journal.uad.ac.id">journal.uad.ac.id</a> Internet Source	6%
2	Awwal, Abdul A. S., Khan M. Iftekharuddin, Mohammad A. Matin, Andrés Márquez, Haipeng Chen, Shuxu Guo, Tian Zhang, and Jian Sun. "The method of evaluation for 1/f low-frequency electronics noise based on compressed sensing", Optics and Photonics for Information Processing VIII, 2014. Publication	3%
3	Submitted to Universiti Teknologi Malaysia Student Paper	1%
4	Submitted to Maryville University Student Paper	1%
5	<a href="http://www.researchgate.net">www.researchgate.net</a> Internet Source	1%
6	Wang, Jian, and Byonghyo Shim. "Exact Recovery of Sparse Signals Using Orthogonal Matching Pursuit: How Many Iterations Do We	1%

Need?", IEEE Transactions on Signal Processing, 2016.

Publication

---

7 Xinpeng Zhang, Niaoqing Hu, Lei Hu, Ling Chen, Zhe Cheng. "A bearing fault diagnosis method based on the low-dimensional compressed vibration signal", Advances in Mechanical Engineering, 2015 1%

Publication

---

8 [documents.mx](http://documents.mx) 1%

Internet Source

---

9 [www.jstage.jst.go.jp](http://www.jstage.jst.go.jp) <1%

Internet Source

---

10 [telkomnika.ee.uad.ac.id](http://telkomnika.ee.uad.ac.id) <1%

Internet Source

---

11 [www.ece.rice.edu](http://www.ece.rice.edu) <1%

Internet Source

---

12 Swastika, Windra, Yoshitada Masuda, Naoko Kawata, Koji Matsumoto, Toshio Suzuki, Ken Iesato, Yuji Tada, Toshihiko Sugiura, Nobuhiro Tanabe, Koichiro Tatsumi, Takashi Ohnishi, and Hideaki Haneishi. "Evaluation of COPD's diaphragm motion extracted from 4D-MRI", Medical Imaging 2015 Image Processing, 2015. <1%

Publication

---

13 [www.dsp.ece.rice.edu](http://www.dsp.ece.rice.edu)

---

Internet Source

<1%

---

14

[control.me.berkeley.edu](http://control.me.berkeley.edu)

Internet Source

<1%

---

15

Sathiya Narayanan, Anamitra Makur. "Camera motion estimation using circulant compressive sensing matrices", 2013 9th International Conference on Information, Communications & Signal Processing, 2013

Publication

<1%

---

16

[www.freepatentsonline.com](http://www.freepatentsonline.com)

Internet Source

<1%

---

17

[personal.its.ac.id](http://personal.its.ac.id)

Internet Source

<1%

---

18

[unstats.un.org](http://unstats.un.org)

Internet Source

<1%

---

19

Emmanuel Candes, Nathaniel Braun, Michael Wakin. "SPARSE SIGNAL AND IMAGE RECOVERY FROM COMPRESSIVE SAMPLES", 2007 4th IEEE International Symposium on Biomedical Imaging: From Nano to Macro, 2007

Publication

<1%

---

20

[mcall.eu.interia.pl](http://mcall.eu.interia.pl)

Internet Source

<1%

---

21 Yair Rivenson. "Single exposure super-resolution compressive imaging by double phase encoding", Optics Express, 07/05/2010 <1%  
Publication

---

22 [www.merl.com](http://www.merl.com) <1%  
Internet Source

---

23 Cognitive Networked Sensing and Big Data, 2014. <1%  
Publication

---

24 "Automata, Languages, and Programming", Springer Nature America, Inc, 2014 <1%  
Publication

---

25 Swastika, Windra, Yoshitada Masuda, Takashi Ohnishi, and Hideaki Haneishi. "Reduction of acquisition time in the intersection profile method for four-dimensional magnetic resonance imaging reconstruction of thoracoabdominal organs", Journal of Medical Imaging, 2015. <1%  
Publication

---

Exclude quotes  On

Exclude matches  Off

Exclude bibliography  On

# Cek Plagiasi Windra Telkomnika 1

---

## GRADEMARK REPORT

---

FINAL GRADE

**/0**

GENERAL COMMENTS

**Instructor**

---

PAGE 1

---

PAGE 2

---

PAGE 3

---

PAGE 4

---

PAGE 5

---

PAGE 6

---

PAGE 7

---

PAGE 8

---